AUG 2 6 2005 K 050/70

## **EXHIBIT 12**

## 510(k) SUMMARY

## Fortoss Vital™

**Applicant** 

Biocomposites Ltd

Keele Science Park

Keele

Staffordshire England ST5 5NL

**Contact Person** 

Mr Simon Fitzer

Tel: +44 1782 338580 Fax +44 1782 338599 Email: sf@biocomposites.com

1. Classification Name:

Synthetic Bone Graft Material

Common/Usual Name:

Synthetic Bone Graft Substitute

Trade/Proprietary Name

Fortoss Vital™ - Bone Graft Substitute

Product Code

LYC

#### **Legally Marketed Predicate Devices**

	Trade Name	<u>Manufacturer</u>	510(k) No
1	Straumann Granules (Bone Graft Substitute)	Straumann AG	K040646
2	Hapset (Bone Graft Material)	Lifecore Biomedical Inc	K910432

## **Device Description**

Fortoss Vital<sup>™</sup> Bone Graft Substitute is a calcium phosphate based bone graft substitute and is provided sterile for single patient use. When Fortoss Vital<sup>™</sup> is placed in the defect, bone grows in apposition to the implant, filling the pores with new bone during the healing process.

Fortoss Vital™ is completely resorbed and replaced with bone during the healing process.

#### Intended Use / Indications

Fortoss Vital™ is indicated for placement in osseous defects to provide a mouldable, resorbable graft in periodontal, maxillofacial and dental implant surgery.

#### Summary of Technology

Fortoss Vital™ is composed of porous calcium salts equivalent to that contained in the predicate devices and to that in routine clinical use. The technologies employed in Fortoss Vital™ and the predicate devices are therefore substantially equivalent. Fortoss Vital™ is presented in granules/powder in the same manner as the predicate devices. The indications, contraindications, risks and potential adverse events are the same and thus substantially equivalent.

### Non Clinical Testing

Non clinical testing has been used to examine the chemical composition of Fortoss Vital™ which satisfies the standard for implantable calcium salts.

#### **Clinical Testing**

Fortoss Vital<sup>™</sup> has been regularly used clinically and no adverse events have been reported in that time concerning the quality, safety or effectiveness of Fortoss Vital<sup>™</sup>.

#### Substantial Equivalence

Documentation provided demonstrates that Fortoss Vital™ is substantially equivalent to the legally marketed predicate devices in design, materials and indications. Fortoss Vital™ is well tolerated and completely incorporated into the defect site into which it is implanted and is safe and effective when used as indicated.



AUG 2 6 2005

Food and Drug Administration 9200 Corporate Boulevard Rockville MD 20850

Biocomposites, Limited. C/O Mr. Simon Fitzer Quality and Regulatory Affairs Manager Keele Science Park Keel Staffordshire, UNITED KINGDOM ST5 5NL

Re: K050170/S002

Trade/Device Name: Fortoss Vital Bone Graft Substitute

Regulation Number: 21 CFR 872.3930 Regulation Name: Bone Grafting Material

Regulatory Class: II Product Code: LPK Dated: August 5, 2005 Received: August 9, 2005

Dear Mr. Fitzer:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <a href="http://www.fda.gov/cdrh/industry/support/index.html">http://www.fda.gov/cdrh/industry/support/index.html</a>.

Sincerely yours,

Chiu Lin, Ph.D.

Director

Division of Anesthesiology, General Hospital Infection Control and Dental Devices Office of Device Evaluation Center for Devices and Radiological Health

Enclosure

**EXHIBIT 1** 

# **INDICATIONS FOR USE**

510(k) Number (if known):				
Device Name:	Fortoss Vital™			
Indications For Use:				
Fortoss Vital ™ is intended mouldable, resorbable graf surgery.	•	seous defects to provide a xillofacial and dental implant		
Prescription Use <u></u> ✓ (Part 21 CFR 801 Subpart	OR D)	Over-The-Counter use(Part 21 CFR 807 Subpart C)		
PLEASE DO NOT WRITE B	ELOW THIS LINE – CO NEEDED	INTINUE ON ANOTHER PAGE IF		
Concurrence of CDRH, Office of Device Evaluation (ODE)				

(Division Sign-Off)
Division of Anesthesiology, General Hospital,
Infection Control, Dental Devices

510(k) Number: KO 50170

Page 1 of 1